PRODUCT SECURITY BULLETIN: PHP VULNERABILITIES

Publication date: 22/10/2026

Last update: 22/10/2026

Manufacturer Name: MicroPort CRM S.r.l.

Address: Via Crescentino s.n.
13040 Saluggia (VC)
Italy

Vulnerability: **CVE-2026-109618**
https://nvd.nist.gov/vuln/detail/cve-2026-109618
CVE-2026-109619
https://nvd.nist.gov/vuln/detail/cve-2026-109619

Software reviewed:

SmartView™ Remote Monitoring Software (RMS)
https://microport-smartview.com/

SyneScope® Easy Cloud
https://synescope.rhythmscope.com/

Following the recently published PHP vulnerabilities, we have reviewed our different web services and applications, as well as the ones delivered by our third-party providers:

- The **SmartView ™ Remote Monitoring Software** does not use PHP in any of its components. It is not affected by these PHP-specific vulnerabilities, so no corrective update is required.
- The **SyneScope® Easy Cloud** application does not use PHP. It is not affected by these PHP-specific vulnerabilities, so no corrective update is required.

No web service provided by MicroPort CardioFlow is vulnerable to the CVE-2026-109618 and CVE-2026-109619.

Third-party providers:

Some of our third-party providers do use PHP. They have either:

- Already deployed the corrective patch to address these vulnerabilities, or
- Have temporarily suspended services using PHP until they can fix the issue. There is **no identified impact** on MicroPort CardioFlow services.